


**2 SETS OF DPA 75 “ BAUR - AUSTRIA”**



## 1. Product Information



The Insulating Oil Tester  
DPA 75

**Measuring the electric breakdown strength of insulating liquids**

**The Insulating Oil Tester DPA 75 is used to measure the electric breakdown strength of insulating liquids (eg insulating oil of transformers etc).**

A judgement criterion for the quality of insulating liquids

The dissipation factor  $\tan \delta$  and the relative permittivity  $\epsilon_r$ , together with the breakdown voltage, are the most important parameter for which insulating liquids are tested.

The breakdown voltage shows the degree of contamination by cellulose fibres and water.

For the operating safety of oil insulated high voltage units the electrical breakdown strength of insulants is the deciding factor.




For further information also see the technical report 'Testing of dielectric properties of insulating liquids' by M. Krüger.  
Obtainable from BAUR under no. 8602e.



**Test vessels and electrodes**

Test vessel 0.4 lt. with lid


IEC 156 Fig I



415-567

Test vessel 0.4 lt. with lid


IEC 156 Fig II



415-566

Test vessel 0.4 lt. with lid


ASTM D 877



415-568

Test vessel 0.4 lt. with lid,  
oil stirrer and setting gauge

ASTM D 1816



415-585



### 5.3 Printer report - standard test



REPORT DPA 75 V1.0  
26.04.94 09:30

REPORT NR.:  
4309

STANDARD:  
IEC 156/63

ELECTRODES : .....

SPACING : .....

TEMP. 20.5 °C  
36.8 °F

S/M 10.6 %


STD. DEV. 7.1

MEAN VALUE 67.2 KV

TEST 6 65.7 KV  
TEST 5 > 75.0 KV  
TEST 4 66.3 KV  
TEST 3 > 75.0 KV  
TEST 2 65.0 KV  
TEST 1 56.0 KV

Reading test results from printer report:  
test data and evaluation results

- > No breakdown happened.
- Breakdown voltage higher than 75 KV.



**Condition:**

If printer is switched off activate it within  
menu INSTRUMENT SETTING.

**CALIBRATION REPORT****Calibrated Equipment**

This calibration reports is belong to following unit identification :

- \* Customer name : R Engineering Support Co., Ltd.
- \* Equipment model : DPA 75
- \* Manufacturer : BAUR
- \* Serial number : 954103078
- \* Report date : February 12, 2008
- \* Due date : February 11, 2009

**Test Equipment**

- \* Equipment model : KA 75
- \* Manufacturer : BAUR
- \* Serial number : 96 42 01 003
- \* Last calibration date : September 06, 2007
- \* Due date : September 05, 2008

As reference, calibrator can be traced back to international and nation standard as attached sheet.

Ambient temperature : 23 °C



C-0802001

Report Service Test:

12.02.2008 16:10:09


=====

Hardware values

	Designation	Unit	Range		Value	
A	U supply	V	10.50 -	14.80	14.44	ok
B	+5V supply	V	4.90 -	5.10	5.00	ok
C	+8V supply	V	8.00 -	8.50	8.14	ok
D	-9V supply	V	-8.90 -	-11.80	-11.56	ok
E	Frequency	Hz	59.70 -	60.30	60.03	ok
F	+40 V counter 0	V	7.00 -	9.00	7.57	ok
G	-40 V counter 0	V	-7.00 -	-9.00	-7.85	ok
H	Amplifier output counter 0	V	0.00 -	0.03	0.00	ok
I	Sine amplifier counter 434	V	1.60 -	1.80	1.69	ok
J	+40 V counter 434	V	22.50 -	25.50	23.20	ok
K	-40 V counter 434	V	-22.50 -	-25.50	-23.99	ok
L	Amplifier output counter 434	V	1.35 -	1.55	1.42	ok
M	Sine amplifier counter 897	V	3.40 -	3.65	3.49	ok
N	+40 V counter 897	V	39.50 -	43.00	40.82	ok
O	-40 V counter 897	V	-39.50 -	-43.00	-41.82	ok
P	Amplifier output counter 897	V	2.85 -	3.15	2.92	ok
Q	Temperature	°C	15.00 -	40.00	29.00	ok

Statistical values

Serial no. : 954103078  
Software version : 1.3  
last calibration : 12.02.108  
Correcting factor kV-value indication : 1.0201  
Correcting factor rate of rise : 1.0277  
Operating hours : 1185  
Operating hours H.V. : 93  
mean breakdown value : 43.47 kV  
No. of tests : 19268  
Supply : Mains operation  
Battery capacity :  
No. of possible single tests :



DPA 75

=====

Seriennummer

Serial no. : 954103078

geprüft mit Kalibrator KA 75 Nr.

tested with calibrator KA 75 no. : 964201003


Anzeige DPA 75 Display	Anzeige KA 75 Display	Differenz (kV)	PC- Auswertung PC- Evaluation
10.05	10.11	-0.06	ok
20.02	20.09	-0.07	ok
30.08	30.06	0.02	ok
39.97	40.00	-0.03	ok
50.02	50.01	0.01	ok
60.08	60.03	0.05	ok
69.96	69.90	0.06	ok
75.03	74.84	0.19	ok

maximale zulässige Abweichung

maximal permissible deviation : ± 1.0 kV

Funktions- u. Sicherheitsprüfung  
nach Prüfanweisung Nr. 471-534 erfüllt

Function and safety check met  
acceptance to test instruction no. 471-534



12.02.2008

Name: *[Signature]*

**SIEMENS****Siemens Limited**

**Industrial Solutions & Services, Calibration Laboratory**  
**79/38 Prime State Project, Soi Pookmit, Srinakarin Rd.,**  
**Nongbon, Pravate, Bangkok 10260 Thailand**  
**Tel. : +66 (0) 2366 0951-8 ext. 201-5 , Fax : +66 (0) 2747 1223**

Certificate No.
<b>501/2007</b>
Number of pages
<b>2</b>

## **Certificate of Calibration**

**Description:** Voltage Calibration Meter

The Certificate of Calibration traceable to the International System of Units maintained at national standards laboratory. This certificate may not be reproduced other than in full, except with the prior written approval of the head of Siemens Calibration Laboratory.

**Manufacturer:** BAUR

The reported uncertainty of measurement is based on standard uncertainty multiplied by a coverage factor K = 2, providing confidence level of approximately 95%.

**Model:** KA 75

**Serial no.:** 964201003

**Inventory no.:** -

The result of Calibration was found accurate as show on date and place of calibration only.

**Customer:** Power Testing Product Co., Ltd.  
 100/33 Moo 5, Bangkruay-Sainoi Road,  
 Bangsithong, Bangkruay  
 Nonthaburi 11130

**Work order no.:** WO-0501/2007

**Procedure used :** Direct measurement with reference standard(s) using in-house calibration procedure(s).

**Reference standards used:**

Description	Model	Manufacturer	Serial no.	Certificate No.	Due Date	Traceability
Multi-Product Calibrator /with coil	5520A	FLUKE	7450203	EL-0075/07	27 Apr 2008	NIST,PTB,BIPM,NPL
Digital Multimeter	3458A	Hewlett Packard	2823A21635	EL-0244/06	09 Nov 2007	PTB,BIPM,NIMT


**Room temperature:** 23 °C ± 2 °C    **Rel. humidity:** 50 % ± 15 %

**Date of Receipt :** 5 Sep 2007

**Date of Calibration :** 6 Sep 2007

**Date of Issue :** 6 Sep 2007

Seal



Head or deputy of the calibration laboratory

Vichit Tipsena



Krisada Nuamnak

**Calibration Results**

Description : Voltage Calibration Meter  
 Manufacturer : BAUR  
 Model / Type : KA 75  
 Serial No. : 964201003  
 Inventory No. : -

**1. Voltage Measurement**

Expected Value	Standard Value	Unit Under Test (UUT)			Estimate Uncertainty(±)
		Value	Error	Tolerance(±)	
15 kV	42.43 µA	14.96 kV	-0.04 kV	0.11 kV	62 V
20 kV	56.57 µA	20.03 kV	0.03 kV	0.14 kV	82 V
30 kV	84.85 µA	30.06 kV	0.06 kV	0.21 kV	0.12 kV
40 kV	113.14 µA	40.07 kV	0.07 kV	0.28 kV	0.16 kV
50 kV	141.42 µA	50.09 kV	0.09 kV	0.35 kV	0.20 kV
60 kV	169.71 µA	60.12 kV	0.12 kV	0.42 kV	0.25 kV
70 kV	197.99 µA	70.15 kV	0.15 kV	0.49 kV	0.29 kV
75 kV	212.13 µA	75.00 kV	0.00 kV	0.53 kV	0.31 kV

**Remark :**

- tested frequency as 60Hz