

Calibration Report



ISSUED BY : ATE CALIBRATION LABORATORY

ADVANCED TECHNOLOGY EQUIPMENT CO.,LTD.
29/19-20-21-22 Rama 9 Road, Huaykwang,
Bangkok 10320, Thailand.

Email : service@advanced.co.th
Tel. 0 - 2643 - 0982 EXT. 117
Fax. 0 - 2246 - 2237

Equipment	Oil Dielectric Test Set	Report No.	51083
Manufacturer	Foster		
Model	OTS 60 AF	CRS No.	N/A
Serial No.	1009 M588990		
ID No.	N/A	Date Received	13-Feb-09

Client R Engineering Support Co.,Ltd
208/32 Muangthong 2/1,Pattanakarn Rd.,
Pravet,Bangkok 10250

Procedures This instrument was allowed to stabilise in the laboratory environment.
The calibration was performed in accordanced with ATE Work Instruction
" PCD 04 "

Ambient Temperature 24.2 °C Relative Humidity 52 %

Date of Calibration 17 - Feb -09

Calibrated By Sutthichai Ponchar

The reported uncertainties are base standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%. The uncertainty relate only to the measured value and do not carry any implication the long term stability of the instrument.

This calibration report may not be reproduced other than in full except with the permission issuing laboratory.

Date of Issue

17-FEB-09



Authorized Signature
Chatree Kruamai

Calibration Report



ISSUED BY : ATE CALIBRATION LABORATORY

ADVANCED TECHNOLOGY EQUIPMENT CO.,LTD.
29/19-20-21-22 Rama 9 Road, Huaykwang,
Bangkok 10320, Thailand.

Email : service@advanced.co.th
Tel. 0 - 2643 - 0982 EXT. 117
Fax. 0 - 2246 - 2237

Traceability and Standards Uses

Item	Description / Model	Serial No.	Calibrated By	Certificate No.	Traceability to	Due Date
1.	Voltage Calibration Meter / VCM 100	020402 / 1621	Megger	COM62497	See Note	26 -Aug - 10

Note : -

NPL : National Physical Laboratory (UK)

Calibration Report



ISSUED BY : ATE CALIBRATION LABORATORY

ADVANCED TECHNOLOGY EQUIPMENT CO.,LTD.
29/19-20-21-22 Rama 9 Road, Huaykwang,
Bangkok 10320, Thailand.

Email : service@advanced.co.th
Tel. 0 - 2643 - 0982 EXT. 117
Fax. 0 - 2246 - 2237

Oil Dielectric Test Set Calibration Data Sheet

Standard Value	UUC. Reading	Deviation	Estimate Uncertainty
9.90 kV.	9 kV.	-0.90 kV.	0.6 kV.
20.1 kV.	19 kV.	-1.10 kV.	0.7 kV.
29.84 kV.	30 kV.	0.16 kV.	0.8 kV.
39.96 kV.	39 kV.	-0.96 kV.	0.9 kV.
50.18 kV.	49 kV.	-1.18 kV.	1.0 kV.
60.65 kV.	59 kV.	-1.65 kV.	1.1 kV.

Note : -

- The uncertainty values stated in this measurement report have been calculated at the 95% level of confidence. This means that the chance that the reported value is in error by more than the stated uncertainty is five in one hundred.