

“Mettler Toledo - Switzerland ”

new arrival

THE EASY ENTRY TO FULLY
AUTOMATIC TITRATION:
METTLER TOLEDO
DL50 Graphix



Simply fully automatic
Neutralization Number Measuring DL50 Graphix

Specific Test Report Prüfnachweis

TR-51'107'010

Instrument
Gerät

Serial number
Seriennummer

Software Version
Software Version

Production date
Herstellungsjahr

Ident number
Identnummer

DL5 0



2.4

2006



including the
mit der

pH-Option



Herewith we certify that the above instrument has been tested and adheres to the technical specifications listed in this document. The complete test report is filed in our archives.

Hiermit bestätigen wir, dass das oben erwähnte Gerät getestet wurde und die in diesem Dokument aufgelisteten technischen Spezifikationen erfüllt. Der komplette Testrapport ist in unseren Archiven abgelegt.

The following tests have been performed according to test instruction VS-51'107'010C, V.2.3, 07.06.00:

Folgende Tests wurden durchgeführt nach Prüfvorschrift VS-51'107'010C, V.2.3, 07.06.00:

	Tested/Adjusted Geprüft/Justiert	Measured Value Messwert	Criterion/Tolerance Kriterium/Toleranz	Passed / Bestanden	
				Yes / Ja	No / Nein
1. 12h Endurance test 12h Dauertest	✓		Operation Funktionsfähigkeit	✓	
2. pH-mV Input pH-mV Eingang					
2.1 0 mV	✓	0.0 mV	0 ± 0.2 mV	✓	
2.2 +1900 mV	✓	1899.8 mV	1900 ± 0.2 mV	✓	
3. Pt100					
3.1 0 °C	✓	0.0 °C	0 ± 0.2 °C	✓	
3.2 130 °C	✓	130 °C	130 ± 0.2 °C	✓	
4. Pt1000					
4.1 0 °C	✓	0.0 °C	0 ± 0.2 °C	✓	
4.2 130 °C	✓	130 °C	130 ± 0.2 °C	✓	
5. Interfaces Schnittstellen	✓		Operation Funktionsfähigkeit	✓	
6. Stirrer Rührer	✓		Operation Funktionsfähigkeit	✓	

Yes / Ja No / Nein

Performed by / Durchgeführt von:

Test criteria have been met
Prüfung bestanden

Schwerzenbach, Date / Datum 3. April 2006

Name / Name H. Buchmeier

Signature / Unterschrift H. B.